

# SOPRANO

## Treatment parameters



## Treatment Settings & Parameters

Selecting the correct settings from the suggested treatment parameters is crucial in delivering a results driven treatment for your client.

The following factors are taken into account when selecting the treatment parameters: **Grid Size**

This determines the size of the treatment area, and dictates the accumulation of energy in an area. In the case of a 100 square cm grid, one would accumulate more energy than when treating a grid of a smaller size, say 50 square centimeters.

The correct grid size must be selected in order to ensure a time-efficient and effective treatment. Prior to starting the treatment, a grid template is used to segment the treatment area, using a white eyeliner pencil, as it has no pigment and will therefore present no target for the laser. If a regular pen was used, the pigment in the ink would present a target and the client could sustain a burn.

Gridding will be demonstrated during your practical sessions.

Example of Gridding on lower legs and facial mapping:



*Example of Gridding on the Face*



**150cm<sup>2</sup>**



**75cm<sup>2</sup>**


**25cm<sup>2</sup>**

**100cm<sup>2</sup>**

**100cm<sup>2</sup>**

**50cm<sup>2</sup>**

**200cm<sup>2</sup>**



**150cm<sup>2</sup>**



